

## **Applied Optoelectronics** in **Medicine**


### Aplikovaná optoelektronika v lékařství

Interdisciplinary course at the CTU Prague (P317APL-E, W, 4 credits)









Optoelectronic sensor concepts for vascular diagnostics – part I
 Optoelektronické koncepty pro vaskulární diagnostiku – část I


© V. Blazek, MedIT, 2017 All rights reserved Lecture 7, Page 1

Scriptum AOM: Applied Optoelectronics in Medicine



### Learning aims of the seventh AOM lecture

- Basics of quantitative Photoplethysmography (PPG) for transcutaneous detection of dermal blood volume changes
- Standardized PPG tests for non-invasive assessment of peripheral venous hemodynamics: a) Muscle pump test and
  - b) Venous occlusion test
- Alternative sensor concepts for blood volume studies in different compartments of human body







© V. Blazek, MedIT, 2017 All rights reserved **Lecture 7**, Page 2



#### Comparison of non-invasive tests of the vascular system with phlebography

A scale of 1-5, in order of ascending quality, has been used

Test	Accuracy								
		igh Insuff.		alf Insuff.	Portability	Training period	Cost	Quantitative results	Туре
Doppler (CW)	1	4	1	4	5	Long	Low	1	Functional
Duplex ultrasound	5	5	4	5	3	Long	High	3	Anatomical & functiona
Plethysmography	5							· · · · · · · · · · · · · · · · · · ·	
segmental Air	4	4	3	4	3	Short	Mediur	n 5	Functional
whole limb	54.57								
standard Photo	1	4	1	4	4	Short	Mediur	3	Functional
calibrated	3	4	2	4	4	SHOIL	Mediaiii	5	runcional
Foot volumetry	3	4	3	4	0	Short	Mediur	n 5	Functional
Impedance	4	2	2	1	3	Short	Mediur	n 4	Functional
Strain gauge	4	2	2	1	3	Short	Mediur	n 4	Functional
Phleborheography	y 4	2	2	1	3	Short	Mediur	n 0	Functional
Phlebography	5	4	4	4	0	Long	High*	2	Anatomical

From: THE NEWS, Phlebology edition, No. 31,k Sept. 1992

\* = invasive

© V. Blazek, MedIT, 2017 All rights reserved Lecture 7, Page 3


Scriptum AOM: Applied Optoelectronics in Medicine



Studies of peripheral hemodynamics using Photoplethysmography (PPG): transcutaneous monitoring of venous and/or arterial blood volume changes in dermal vascular plexus.

Milestones/Evolution steps:

- 1) 1938: first PPG system, created by Hertzman
- 2) 1979: LRR, first full portable PPG system
- 3) 1989: first quantitative, PC controlled PPG system with calibration software for each measurement
- 4) and today?
  - "intelligent" OES strategies like Smart PPG
  - camera based PPG Imager for spatial & time resolved skin perfusion studies










© V. Blazek, MedIT, 2017 All rights reserved Lecture 7, Page 4



#### Discovery of the "optical windows" through the skin in the 70ties I.c. at RWTH

allowed the construction of advanced optoelectronic sensors (optrodes) for non-invasive measurements of venous and/or arterial blood volume changes




© V. Blazek, MedIT, 2017 All rights reserved Lecture 7, Page 5

Scriptum AOM: Applied Optoelectronics in Medicine




### Fundamentals of Photoplethysmography (PPG) - one of the world wide well used methods for blood volume monitoring in dermal vascular network



Schematic visualisation of the correlation between the PPG signal and the blood volume changes in transilluminated vascular network under the sensor



Fundamentals of Photoplethysmography (PPG) - one of the world wide well used methods for blood volume monitoring in dermal vascular network




© V. Blazek, MedIT, 2017 All rights reserved Lecture 7, Page 7

Scriptum AOM: Applied Optoelectronics in Medicine



Fundamentals of Photoplethysmography (PPG) - one of the world wide well used methods for blood volume monitoring in dermal vascular network




Photon distribution in the skin and an estimation of the signal contents detected by the PPG sensor (T: tissue, V: venous signal, A: arterial signal). From 10<sup>6</sup> Photons injected to the skin in this model scenario only 50 were detected (-43 dB)

© V. Blazek, MedIT, 2017 All rights reserved Lecture 7, Page 8



Fundamentals of Photoplethysmography (PPG) - one of the world wide well used methods for blood volume monitoring in dermal vascular network



Time dependent visualisation of the typical PPG signal components

© V. Blazek, MedIT, 2017 All rights reserved **Lecture 7**, Page 9


Scriptum AOM: Applied Optoelectronics in Medicine



### PPG measuring system "for beginners"

Optoelectronic components in the sensor


- "Arterial sensor": LD 242 as light source and BPW (photodiode) as light detector; both without focussing lenses.
- "Venous sensor": TIL31 as light source and TIL81 (phototransistor) as light detector; both with focussing lenses.



Possible sensor interfaces



### PPG measuring system "for beginners": schema of the electrical part




© V. Blazek, MedIT, 2017 All rights reserved **Lecture 7**, Page 11

Scriptum AOM: Applied Optoelectronics in Medicine




### Fundamentals of Photoplethysmography (PPG) - one of the world wide well used methods for blood volume monitoring in dermal vascular network



Block diagram of an  $\mu$ C aided PPG system (so called quantitative Photoplethysmography). In a closed loop the intensity of illumination will be adjusted to compensate individual optical transfer function of the skin.

RWTHAACHEN UNIVERSITY

#### Functional monitoring of vascular hemodynamics - standardized clinical PPG tests




© V. Blazek, MedIT, 2017 All rights reserved Lecture 7, Page 13

Scriptum AOM: Applied Optoelectronics in Medicine



#### First and most used PPG test:

MPT for functional assessment of the global blood transport properties of the leg vein system during standardized exercise and for evaluating the efficiency of the calf pump







### Wie funktioniert der venöse Rückstrom des Blutes beim Menschen?




Venenklappen



Muskel-Venen-Pumpe



Thorax-Pumpe




© V. Blazek, MedIT, 2017 All rights reserved Lecture 7, Page 15

Scriptum AOM: Applied Optoelectronics in Medicine








© V. Blazek, MedIT, 2017 All rights reserved Lecture 7, Page 16






© V. Blazek, MedIT, 2017 All rights reserved Lecture 7, Page 17

Scriptum AOM: Applied Optoelectronics in Medicine




### Zur Häufigkeit der Venenleiden...



© V. Blazek, MedIT, 2017 All rights reserved Lecture 7, Page 18






© V. Blazek, MedIT, 2017 All rights reserved Lecture 7, Page 19

Scriptum AOM: Applied Optoelectronics in Medicine



### Klinisch eingeführte PPG-Geräte für funktionelle Venendiagnostik






© V. Blazek, MedIT, 2017 All rights reserved Lecture 7, Page 20



#### Weitere Vorteile der quantitativen Photoplethysmographen

Rechnergesteuerter Messablauf, einfache Bedienung, Datenspeicherung und -analyse





Hauptmenü des D-PPG-Gerätes


© V. Blazek, MedIT, 2017 All rights reserved Lecture 7, Page 21


Scriptum AOM: Applied Optoelectronics in Medicine



Selbsttest und automatische Signalkalibrierung durch Regelung des Sendestroms des PPG-Sensors (succesive 8bit-Approximation).

Beispiel eines PPG-MPT-Protokolls: Messwiderholung an der gleichen Hautstelle, jedoch mit einem Graufilter zwischen Sensor und Haut






© V. Blazek, MedIT, 2017 All rights reserved Lecture 7, Page 22



### PPG muscle pump test





#### World wide standardized PPG degrees:


Normal : To > 25 sec
PPG stage I : To 24 to 20 sec
PPG stage II : To 19 to 10 sec
PPG stage III : To < 10 sec


© V. Blazek, MedIT, 2017 All rights reserved **Lecture 7**, Page 23

Scriptum AOM: Applied Optoelectronics in Medicine



### MPT parameter venous refilling time $\rm T_0$ and venous drainage $\rm V_0$ in relation to clinical severity of CVI $^*$





 $^*\, Blazek, V., Schultz-Ehrenburg, U.: Quantitative Photoplethysmography. \, VDI \, Verlag \, 1996$ 



Quantitative Photoplethysmography: venous muscle pump test Tourniquet test before invasive therapy (sclerotherapy or surgery)












© V. Blazek, MedIT, 2017 All rights reserved Lecture 7, Page 25

Scriptum AOM: Applied Optoelectronics in Medicine




Quantitative Photoplethysmography: venous muscle pump test Tourniquet test before invasive therapy (sclerotherapy or surgery)



Patient with improvable CVI. Measurements without (left) ant with tourniquet (right)



Example: Multi-wavelength venous recordings using "MedIT smart PPG"




© V. Blazek, MedIT, 2017 All rights reserved Lecture 7, Page 27

Scriptum AOM: Applied Optoelectronics in Medicine



### PPG and invasive vein pressure measurement in comparison: Experimental setup and patient/sensor position during the study






© V. Blazek, MedIT, 2017 All rights reserved Lecture 7, Page 28

### PPG and invasive vein pressure measurement in comparison:

Typical PPG (non-invasive) and VPM (invasive) records, measured simultaneously




© V. Blazek, MedIT, 2017 All rights reserved Lecture 7, Page 29

Scriptum AOM: Applied Optoelectronics in Medicine



#### PPG and invasive vein pressure measurement in comparison:

Study results from 5 different research groups




© V. Blazek, MedIT, 2017 All rights reserved Lecture 7, Page 30



#### Second mostly used PPG test:

Venous occlusion test (VOT) for quantitative assessment of venous outflow dynamics




© V. Blazek, MedIT, 2017 All rights reserved Lecture 7, Page 31

Scriptum AOM: Applied Optoelectronics in Medicine




#### Second mostly used PPG test: VOT for the quantitative assessment of venous outflow





Scriptum AOM: Applied Optoelectronics in Medicine

RWTHAACHEN UNIVERSITY



© V. Blazek, MedIT, 2017 All rights reserved **Lecture 7**, Page 33


Scriptum AOM: Applied Optoelectronics in Medicine



### Alternative sensor concepts for blood volume studies in different compartments of human body:

1) Strain gauge Plethysmography \*

Um die untersuchte Extremität wird ein dünner, mit leitendem Fluid gefüllter und hoch dehnbarer Schlauch gelegt, dessen Längenveränderung als Änderung des elektrischen Widerstandes des Fluidmaterials registriert wird.



\* WHITNEY, R.J.: The measurement of changes in human limb volume by means of a mercury-in-rubber strain gauge. J. Physiol. 109, (1949), 5ff

© V. Blazek, MedIT, 2017 All rights reserved Lecture 7, Page 34



### Alternative sensor concepts for blood volume studies in different compartments of human body:

1) Strain gauge Plethysmography \*





© V. Blazek, MedIT, 2017 All rights reserved Lecture 7, Page 35

Scriptum AOM: Applied Optoelectronics in Medicine



### Alternative sensor concepts for blood volume studies in different compartments of human body:

1) Strain gauge Plethysmography \*



```
Digitale
Strain-gauge-
Plethysmographie

Untersuchungsprotokoll
lfd. Daten-Nr. 00006
(Version 1.51)

Name, Vorname
Geburtsdatum
Extremität
Datum, Unrzeit
Klinische Diagnose:
Okkluston/Bemerkung:

SGP-Kurve:

0 30 60 90 120 130 140 s

Arterieller Einstrom: AF = 5.2 %/min
Venöse Kapazität : VC = 3.6 %
Abstromkapazität: 0C = 3.3 %
3s - Abstromkapazität: 03 = 2.2 %
5e - Abstromkapazität: 05 = 2.4 %
Venöser Abstrom : VC = 38.0 %/min
Venöse Halbwertszeit : TH = 2.5 s
```


© V. Blazek, MedIT, 2017 All rights reserved Lecture 7, Page 36



Alternative sensor concepts for blood volume studies in different compartments of human body:
2) Fiberoptic Plethysmography (RWTH Aachen)

Optical attenuation of FO banding sensor:

$$\alpha_{LWL} = f(r, \lambda, n_{Core}, n_{Coat})$$






© V. Blazek, MedIT, 2017 All rights reserved **Lecture 7**, Page 37

Scriptum AOM: Applied Optoelectronics in Medicine



### Alternative sensor concepts for blood volume studies in different compartments of human body:

2) Fiberoptic Plethysmography (RWTH Aachen)



© V. Blazek, MedIT, 2017 All rights reserved Lecture 7, Page 38



### Alternative sensor concepts for blood volume studies in different compartments of human body:3) Venous Impedance Plethysmography (RWTH Aachen)

**Methodology:** Two "injection" electrodes are connected to the AC source (20-70 kHz; ca. 1mA), two "sensing" electrodes detect changes of tissue impedance as a function of total blood volume in the assessed extremity segment \*





\* After WHEELER et al., 1971

© V. Blazek, MedIT, 2017 All rights reserved **Lecture 7**, Page 39

Scriptum AOM: Applied Optoelectronics in Medicine



### Alternative sensor concepts for blood volume studies in different compartments of human body:

- Air-Plethysmography (NICOLAIDES, IC London)
- Water-Plethysmography (PARTSCH, Univ. of Vienna)





© V. Blazek, MedIT, 2017 All rights reserved Lecture 7, Page 40



### Alternative sensor concepts for blood volume studies in different compartments of human body:

- 6) Gravimetric Plethysmography (RWTH Aachen)
- 7) Microwave Plethysmography (RWTH Aachen)





© V. Blazek, MedIT, 2017 All rights reserved Lecture 7, Page 41

Scriptum AOM: Applied Optoelectronics in Medicine



### Alternative sensor concepts for blood volume studies in different compartments of human body:

8) Compression Interface Plethysmography (RWTH Aachen) with a novel *calibrated tandem sensor* concept (Inflator with thin and flexible pneumatic sensor and piezo-electronic pressure converter), which avoids conceptional limitations of prevailing measuring systems.



Our prototype system data:

thin, flexible sensor
14x8x4 cm size 245 gram incl. 9V accu
analogue & digital data output
measuring range 0-100 mmHg
high time resolution
high sensitivity
low cost concept
PC/laptop compatible

© V. Blazek, MedIT, 2017 All rights reserved Lecture 7, Page 42 electronics in Medicine

### Alternative sensor concepts for blood volume studies in different compartments of human body:

8) Compression Interface Plethysmography (RWTH Aachen)

Typical results: Pressure behavior during standardized MP





Resting pressure:

Resting pressure:

Po = 23.8 mmHg

Workling pressure:

Po = 23.8 mmHg

Workling pressure:

Po = 23.8 mmHg

Resting pressure:  $p_0 = 23.8 \text{ mmHg}$ Working pressure:  $p_A = 15.4 \text{ mmHg}$ Venous drainage:  $p_A = 15.4 \text{ mmHg}$ Venous refilling time:  $p_A = 15.9 \text{ sec}$ 

© V. Blazek, MedIT, 2017 All rights reserved Lecture 7, Page 43

Scriptum AOM: Applied Optoelectronics in Medicine



Citát pro sedmou přednášku / Quotation of the lecture 7:

# "Study the past if you would diving the future"



CONFUCIUS (551 - 479 B.C.) Chinese philosopher

© V. Blazek, MedIT, 2017 All rights reserved Lecture 7, Page 44

